



## **ADDENDUM NO. 1**

**October 19, 2021**

**BID # 22-003 – Chemistry & Physics – Light Microscope**

**TO: ALL POTENTIAL BIDDERS**

**FROM: Vanessa Maulden**

**Troy University**

**Purchasing & Asset Management Dept.**

---

Dear potential vendor,

Please see the following questions that we have received regarding this bid.

1. Is transmitted or reflected light needed, or both?  
Both are needed.
2. Is quantified polarized light needed?  
Polarized light is mainly used for the differential interference contrast (DIC). If DIC can be performed, it doesn't have to be quantified.

**END OF ADDENDUM NO. 1**